

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
 Data File : VV025237.D
 Acq On : 27 Mar 2022 06:52
 Operator : SY/MD
 Sample : N2110-09
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0PC2

Quant Time: Mar 28 03:49:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 03:13:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	146108	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	139350	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60573	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	58028	4.252	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.000%	
7) Chloroethane-d5	1.571	69	50967	4.450	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.000%	
11) 1,1-Dichloroethene-d2	2.111	63	92913	3.143	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.800%	
20) 2-Butanone-d5	3.908	46	61987	45.596	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	91.200%	
24) Chloroform-d	4.349	84	85952	4.826	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.600%	
26) 1,2-Dichloroethane-d4	5.037	65	39188	4.978	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	99.600%	
32) Benzene-d6	5.050	84	174487	4.949	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.000%	
36) 1,2-Dichloropropane-d6	6.069	67	50414	5.103	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.000%	
41) Toluene-d8	7.317	98	147933	4.575	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	91.400%	
43) trans-1,3-Dichloroprop...	7.625	79	13859	4.300	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.000%	
46) 2-Hexanone-d5	8.091	63	62630	53.802	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	107.600%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	30012	4.556	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	91.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	49816	5.121	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.400%	
Target Compounds						Qvalue
12) 1,1-Dichloroethene	2.124	96	3183	0.216	ug/L	# 1
17) Methyl tert-butyl Ether	2.773	73	22722	1.221	ug/L	97
18) trans-1,2-Dichloroethene	2.770	96	477	0.045	ug/L	76
22) cis-1,2-Dichloroethene	3.918	96	13645	1.272	ug/L	97
25) Chloroform	4.381	83	5924	0.303	ug/L	90
34) Trichloroethene	5.918	95	15140	1.349	ug/L	93
47) Tetrachloroethene	7.982	164	7226909	873.210	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025237.D
Acq On : 27 Mar 2022 06:52
Operator : SY/MD
Sample : N2110-09
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0PC2

Quant Time: Mar 28 03:49:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 03:13:34 2022
Response via : Initial Calibration

