

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032622\
 Data File : VW025212.D
 Acq On : 26 Mar 2022 20:43
 Operator : SY/MD
 Sample : N2097-06
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD5

Quant Time: Mar 28 02:04:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Mon Mar 28 01:44:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	151847	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	145987	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62793	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	65249	4.600	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.568	69	57760	4.852	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.111	63	102359	3.332	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.600%
20) 2-Butanone-d5	3.908	46	70307	49.762	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.520%
24) Chloroform-d	4.349	84	91903	4.965	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.037	65	42362	5.178	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.600%
32) Benzene-d6	5.050	84	190774	5.165	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.400%
36) 1,2-Dichloropropane-d6	6.069	67	53356	5.155	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.000%
41) Toluene-d8	7.317	98	162581	4.799	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.000%
43) trans-1,3-Dichloroprop...	7.625	79	15411	4.564	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	91.200%
46) 2-Hexanone-d5	8.092	63	57213	46.914	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.820%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	35214	5.102	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	53859	5.340	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	106.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032622\
Data File : VV025212.D
Acq On : 26 Mar 2022 20:43
Operator : SY/MD
Sample : N2097-06
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD5

Quant Time: Mar 28 02:04:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Mon Mar 28 01:44:37 2022
Response via : Initial Calibration

