

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032724\
 Data File : VV034855.D
 Acq On : 27 Mar 2024 11:18
 Operator : SY/MD
 Sample : VV0327WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 28 01:54:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 28 01:53:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	269122	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	258835	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	122869	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	127506	49.059	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	98.120%
7) Chloroethane-d5	1.532	69	123001	54.487	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	108.980%
11) 1,1-Dichloroethene-d2	2.060	65	55924	49.086	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.180%
21) 2-Butanone-d5	3.786	46	173648	112.653	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	112.650%
24) Chloroform-d	4.246	84	232700	53.251	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	106.500%
26) 1,2-Dichloroethane-d4	4.937	65	157056	56.609	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	113.220%
32) Benzene-d6	4.957	84	456773	52.537	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	105.080%
36) 1,2-Dichloropropane-d6	5.986	67	150004	54.332	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	108.660%
41) Toluene-d8	7.243	98	388596	49.629	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.260%
43) trans-1,3-Dichloroprop...	7.551	79	58841	43.839	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.680%
47) 2-Hexanone-d5	8.024	63	102211	97.918	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.920%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	208818	54.924	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	109.840%
66) 1,2-Dichlorobenzene-d4	11.558	152	144945	52.541	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032724\
Data File : VV034855.D
Acq On : 27 Mar 2024 11:18
Operator : SY/MD
Sample : VV0327WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 28 01:54:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 28 01:53:40 2024
Response via : Initial Calibration

