

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032724\
 Data File : VW034856.D
 Acq On : 27 Mar 2024 11:42
 Operator : SY/MD
 Sample : P1851-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Quant Time: Mar 28 01:54:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 28 01:53:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	149047	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	146817	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	69348	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	127401	88.508	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 177.020%#		
7) Chloroethane-d5	1.532	69	118522	94.800	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 189.600%#		
11) 1,1-Dichloroethene-d2	2.060	65	58736	93.088	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 186.180%#		
21) 2-Butanone-d5	3.786	46	183550	215.007	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 215.010%#		
24) Chloroform-d	4.246	84	258152	106.668	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 213.340%#		
26) 1,2-Dichloroethane-d4	4.941	65	170275	110.818	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 221.640%#		
32) Benzene-d6	4.960	84	502282	101.849	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 203.700%#		
36) 1,2-Dichloropropane-d6	5.986	67	170587	108.929	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 217.860%#		
41) Toluene-d8	7.243	98	425363	95.773	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 191.540%#		
43) trans-1,3-Dichloroprop...	7.551	79	73569	96.633	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 193.260%#		
47) 2-Hexanone-d5	8.024	63	113555	191.786	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 191.790%#		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	246001	114.071	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 228.140%#		
66) 1,2-Dichlorobenzene-d4	11.561	152	167537	107.601	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 215.200%#		

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032724\
Data File : VV034856.D
Acq On : 27 Mar 2024 11:42
Operator : SY/MD
Sample : P1851-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Mar 28 01:54:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 28 01:53:40 2024
Response via : Initial Calibration

