

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032724\
 Data File : VV034860.D
 Acq On : 27 Mar 2024 13:17
 Operator : SY/MD
 Sample : PB159798TB 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Quant Time: Mar 28 01:55:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Mar 28 01:53:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	363481	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	353878	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	167681	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	107361	30.584	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	61.160%
7) Chloroethane-d5	1.526	69	78359	25.700	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	51.400%#
11) 1,1-Dichloroethene-d2	2.056	65	51139	33.234	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.460%
21) 2-Butanone-d5	3.796	46	167145	80.285	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.280%
24) Chloroform-d	4.246	84	230610	39.073	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.140%
26) 1,2-Dichloroethane-d4	4.940	65	149349	39.857	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	79.720%
32) Benzene-d6	4.957	84	440936	37.094	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	74.180%
36) 1,2-Dichloropropane-d6	5.985	67	149340	39.564	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	79.120%
41) Toluene-d8	7.243	98	373176	34.859	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	69.720%#
43) trans-1,3-Dichloroprop...	7.551	79	67299	36.674	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.340%
47) 2-Hexanone-d5	8.027	63	120972	84.765	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.770%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	219512	42.230	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.460%
66) 1,2-Dichlorobenzene-d4	11.561	152	149241	39.641	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.280%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032724\
Data File : VV034860.D
Acq On : 27 Mar 2024 13:17
Operator : SY/MD
Sample : PB159798TB 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Mar 28 01:55:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Mar 28 01:53:40 2024
Response via : Initial Calibration

