

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW032822\  
 Data File : VW025253.D  
 Acq On : 28 Mar 2022 12:02  
 Operator : SY/MD  
 Sample : N2097-21  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 6 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Mar 29 03:28:00 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR032622WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Tue Mar 29 03:26:38 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114   | 144931   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850  | 117   | 139230   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.246 | 152   | 58270    | 5.000    | ug/L  | 0.00     |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.304  | 65    | 62975    | 4.652    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 93.000%  |
| 7) Chloroethane-d5            | 1.568  | 69    | 57187    | 5.033    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 100.600% |
| 11) 1,1-Dichloroethene-d2     | 2.108  | 63    | 100763   | 3.436    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 68.800%  |
| 20) 2-Butanone-d5             | 3.921  | 46    | 52804    | 39.157   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 78.320%  |
| 24) Chloroform-d              | 4.349  | 84    | 87534    | 4.955    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 99.000%  |
| 26) 1,2-Dichloroethane-d4     | 5.034  | 65    | 40454    | 5.180    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 103.600% |
| 32) Benzene-d6                | 5.050  | 84    | 175186   | 4.974    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 99.400%  |
| 36) 1,2-Dichloropropane-d6    | 6.069  | 67    | 50581    | 5.124    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 102.400% |
| 41) Toluene-d8                | 7.317  | 98    | 149027   | 4.613    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 92.200%  |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 14775    | 4.588    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 91.800%  |
| 46) 2-Hexanone-d5             | 8.092  | 63    | 56096    | 48.230   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 96.460%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 33004    | 5.014    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 100.200% |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 50503    | 5.396    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 108.000% |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV032822\  
Data File : VV025253.D  
Acq On : 28 Mar 2022 12:02  
Operator : SY/MD  
Sample : N2097-21  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 6 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VHBLK001

Quant Time: Mar 29 03:28:00 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR032622WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Tue Mar 29 03:26:38 2022  
Response via : Initial Calibration

