

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032822\
 Data File : VW025266.D
 Acq On : 28 Mar 2022 17:08
 Operator : SY/MD
 Sample : N2101-23
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAZ69

Quant Time: Mar 29 03:30:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 29 03:26:38 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	143552	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	139873	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58144	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	58652	4.374	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	87.400%
7) Chloroethane-d5	1.568	69	54615	4.853	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.108	63	95165	3.276	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.600%
20) 2-Butanone-d5	3.918	46	58731	43.971	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.940%
24) Chloroform-d	4.349	84	85299	4.874	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
26) 1,2-Dichloroethane-d4	5.034	65	39871	5.155	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.000%
32) Benzene-d6	5.050	84	167435	4.732	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.600%
36) 1,2-Dichloropropane-d6	6.069	67	47901	4.830	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.317	98	143546	4.423	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
43) trans-1,3-Dichloroprop...	7.625	79	14393	4.449	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.000%
46) 2-Hexanone-d5	8.091	63	52020	44.520	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	89.040%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	30517	4.615	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	49104	5.258	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032822\
Data File : VV025266.D
Acq On : 28 Mar 2022 17:08
Operator : SY/MD
Sample : N2101-23
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAZ69

Quant Time: Mar 29 03:30:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 29 03:26:38 2022
Response via : Initial Calibration

