

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032822\
 Data File : VV025275.D
 Acq On : 28 Mar 2022 21:01
 Operator : SY/MD
 Sample : VV0328WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK328

Quant Time: Mar 29 04:58:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 29 04:58:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	140003	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	133062	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56394	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	56119	4.291	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	85.800%	
7) Chloroethane-d5	1.571	69	52415	4.775	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	95.600%	
11) 1,1-Dichloroethene-d2	2.111	63	88205	3.114	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.200%	
20) 2-Butanone-d5	3.902	46	65427	50.225	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	100.460%	
24) Chloroform-d	4.349	84	87677	5.137	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.800%	
26) 1,2-Dichloroethane-d4	5.034	65	41557	5.509	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	110.200%	
32) Benzene-d6	5.050	84	174190	5.174	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	103.400%	
36) 1,2-Dichloropropane-d6	6.069	67	51185	5.425	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	108.600%	
41) Toluene-d8	7.317	98	150559	4.876	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.600%	
43) trans-1,3-Dichloroprop...	7.625	79	14282	4.640	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	92.800%	
46) 2-Hexanone-d5	8.091	63	54338	48.884	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	97.760%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32668	5.193	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	103.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	51139	5.646	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	113.000%	
Target Compounds						
25) Chloroform	4.384	83	2376	0.127	ug/L	Qvalue 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032822\
Data File : VV025275.D
Acq On : 28 Mar 2022 21:01
Operator : SY/MD
Sample : VV0328WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK328

Quant Time: Mar 29 04:58:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 29 04:58:30 2022
Response via : Initial Calibration

