

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032822\
 Data File : VW025279.D
 Acq On : 28 Mar 2022 22:35
 Operator : SY/MD
 Sample : N2101-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAZ08

Quant Time: Mar 29 05:16:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 29 04:58:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	139091	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	138639	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	57617	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	47917	3.688	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.800%
7) Chloroethane-d5	1.571	69	48268	4.426	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	88.600%
11) 1,1-Dichloroethene-d2	2.111	63	76476	2.717	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	54.400%#
20) 2-Butanone-d5	3.892	46	68859	53.207	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.420%
24) Chloroform-d	4.352	84	83592	4.930	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.600%
26) 1,2-Dichloroethane-d4	5.034	65	39513	5.272	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.400%
32) Benzene-d6	5.053	84	160331	4.571	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.072	67	48500	4.934	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	98.600%
41) Toluene-d8	7.316	98	137842	4.285	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	85.600%
43) trans-1,3-Dichloroprop...	7.625	79	13160	4.104	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.000%
46) 2-Hexanone-d5	8.088	63	53501	46.195	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.400%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33285	5.078	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	48933	5.288	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.800%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032822\
Data File : VV025279.D
Acq On : 28 Mar 2022 22:35
Operator : SY/MD
Sample : N2101-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAZ08

Quant Time: Mar 29 05:16:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 29 04:58:30 2022
Response via : Initial Calibration

