

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032822\
 Data File : VW025281.D
 Acq On : 28 Mar 2022 23:22
 Operator : SY/MD
 Sample : N2101-19
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAZ54

Quant Time: Mar 29 05:16:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Mar 29 04:58:30 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	139701	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	135684	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58132	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	47422	3.634	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	72.600%	
7) Chloroethane-d5	1.571	69	50646	4.624	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	92.400%	
11) 1,1-Dichloroethene-d2	2.111	63	76837	2.718	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	54.400%#	
20) 2-Butanone-d5	3.889	46	70091	53.922	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	107.840%	
24) Chloroform-d	4.352	84	82799	4.862	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.200%	
26) 1,2-Dichloroethane-d4	5.034	65	38631	5.132	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.600%	
32) Benzene-d6	5.050	84	164780	4.800	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.000%	
36) 1,2-Dichloropropane-d6	6.069	67	49130	5.107	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	102.200%	
41) Toluene-d8	7.317	98	139605	4.434	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.600%	
43) trans-1,3-Dichloroprop...	7.625	79	12860	4.098	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	82.000%	
46) 2-Hexanone-d5	8.088	63	55763	49.197	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.400%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33989	5.299	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	48865	5.234	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032822\
Data File : VV025281.D
Acq On : 28 Mar 2022 23:22
Operator : SY/MD
Sample : N2101-19
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAZ54

Quant Time: Mar 29 05:16:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Mar 29 04:58:30 2022
Response via : Initial Calibration

