

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032824\
 Data File : VW034882.D
 Acq On : 28 Mar 2024 13:35
 Operator : SY/MD
 Sample : P1852-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 29 00:24:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Mar 29 00:23:40 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	349588	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	344180	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	160494	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	147023	43.547	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.100%
7) Chloroethane-d5	1.532	69	132814	45.292	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.580%
11) 1,1-Dichloroethene-d2	2.059	65	64366	43.492	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.980%
21) 2-Butanone-d5	3.793	46	153593	76.707	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	76.710%
24) Chloroform-d	4.249	84	235946	41.566	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.140%
26) 1,2-Dichloroethane-d4	4.940	65	151862	42.138	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.280%
32) Benzene-d6	4.960	84	486363	42.069	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.140%
36) 1,2-Dichloropropane-d6	5.989	67	155867	42.456	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	84.920%
41) Toluene-d8	7.243	98	427661	41.075	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.140%
43) trans-1,3-Dichloroprop...	7.554	79	67430	37.781	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.560%
47) 2-Hexanone-d5	8.027	63	98442	70.922	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	70.920%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	203748	40.302	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.600%
66) 1,2-Dichlorobenzene-d4	11.561	152	154451	42.862	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.720%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032824\
Data File : VV034882.D
Acq On : 28 Mar 2024 13:35
Operator : SY/MD
Sample : P1852-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 29 00:24:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Mar 29 00:23:40 2024
Response via : Initial Calibration

