

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW032922\
 Data File : VW025301.D
 Acq On : 29 Mar 2022 18:24
 Operator : SY/MD
 Sample : N2110-14
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Mar 30 02:07:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Mar 30 02:04:47 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	141666	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	141162	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	58494	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	67449	5.097	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 102.000%		
7) Chloroethane-d5	1.568	69	59479	5.355	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 107.200%		
11) 1,1-Dichloroethene-d2	2.111	63	98536	3.438	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 68.800%		
20) 2-Butanone-d5	3.886	46	59837	45.395	ug/L	-0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	= 90.800%		
24) Chloroform-d	4.349	84	83054	4.809	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 96.200%		
26) 1,2-Dichloroethane-d4	5.034	65	37581	4.923	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 98.400%		
32) Benzene-d6	5.050	84	168806	4.727	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 94.600%		
36) 1,2-Dichloropropane-d6	6.069	67	48540	4.850	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 97.000%		
41) Toluene-d8	7.317	98	145939	4.455	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 89.200%		
43) trans-1,3-Dichloroprop...	7.625	79	14140	4.331	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 86.600%		
46) 2-Hexanone-d5	8.088	63	46545	39.471	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 78.940%		
56) 1,1,2,2-Tetrachloroeth...	10.214	84	29082	4.358	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 87.200%		
66) 1,2-Dichlorobenzene-d4	11.622	152	47579	5.064	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 101.200%		
Target Compounds						Qvalue
25) Chloroform	4.375	83	2423	0.128	ug/L	85
47) Tetrachloroethene	7.979	164	2476	0.295	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV032922\
Data File : VV025301.D
Acq On : 29 Mar 2022 18:24
Operator : SY/MD
Sample : N2110-14
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Mar 30 02:07:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Mar 30 02:04:47 2022
Response via : Initial Calibration

