

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033022\
 Data File : VV025307.D
 Acq On : 30 Mar 2022 11:01
 Operator : SY/MD
 Sample : VV0330WBL01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK330

Quant Time: Mar 31 01:26:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 31 01:26:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.612	114	144883	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.847	117	137237	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	56827	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	69250	5.117	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.400%
7) Chloroethane-d5	1.568	69	56246	4.952	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	99.000%
11) 1,1-Dichloroethene-d2	2.108	63	102279	3.489	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.800%
20) 2-Butanone-d5	3.896	46	41831	31.030	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	62.060%
24) Chloroform-d	4.342	84	85822	4.859	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.200%
26) 1,2-Dichloroethane-d4	5.027	65	38338	4.911	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.200%
32) Benzene-d6	5.043	84	172267	4.962	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
36) 1,2-Dichloropropane-d6	6.066	67	47624	4.894	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.800%
41) Toluene-d8	7.310	98	147434	4.630	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
43) trans-1,3-Dichloroprop...	7.622	79	14668	4.621	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.400%
46) 2-Hexanone-d5	8.088	63	43528	37.968	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.940%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	29546	4.554	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	47148	5.166	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033022\
Data File : VV025307.D
Acq On : 30 Mar 2022 11:01
Operator : SY/MD
Sample : VV0330WBL01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK330

Quant Time: Mar 31 01:26:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 31 01:26:10 2022
Response via : Initial Calibration

