

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW033022\
 Data File : VW025308.D
 Acq On : 30 Mar 2022 11:39
 Operator : SY/MD
 Sample : N2165-07
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6Q6

Quant Time: Mar 31 01:26:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Mar 31 01:26:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.613	114	145699	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	140058	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	57882	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	68989	5.069	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	101.400%
7) Chloroethane-d5	1.564	69	54842	4.801	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.000%
11) 1,1-Dichloroethene-d2	2.105	63	98213	3.332	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.600%
20) 2-Butanone-d5	3.896	46	42629	31.445	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	62.900%
24) Chloroform-d	4.339	84	83264	4.688	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.800%
26) 1,2-Dichloroethane-d4	5.027	65	35639	4.540	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.800%
32) Benzene-d6	5.043	84	165750	4.678	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.600%
36) 1,2-Dichloropropane-d6	6.063	67	46502	4.683	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.600%
41) Toluene-d8	7.310	98	142063	4.371	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.400%
43) trans-1,3-Dichloroprop...	7.619	79	13638	4.210	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.200%
46) 2-Hexanone-d5	8.088	63	42687	36.484	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	72.960%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	27730	4.188	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	45682	4.914	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%
Target Compounds						
16) Methylene chloride	2.503	84	7106	0.617	ug/L	Qvalue 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033022\
Data File : VV025308.D
Acq On : 30 Mar 2022 11:39
Operator : SY/MD
Sample : N2165-07
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6Q6


Quant Time: Mar 31 01:26:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Mar 31 01:26:10 2022
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.617 ug/L
RT: 2.503 min Scan# 4
Delta R.T. -0.003 min
Lab File: VV025308.D
Acq: 30 Mar 2022 11:39

Instrument :
MSVOA_V
ClientSampleId :
EW6Q6



Tgt Ion: 84 Resp: 7106

Ion	Ratio	Lower	Upper
84	100		
86	61.1	43.5	80.7
49	91.4	74.5	138.3

