



Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV033122\
 Data File : VV025324.D
 Acq On : 31 Mar 2022 09:41
 Operator : SY/MD
 Sample : BFB292
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFB292

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Title : TRACE VOA SFAM1.0
 Last Update : Sat Apr 02 00:37:35 2022



AutoFind: Scans 2808, 2809, 2810; Background Corrected with Scan 2797

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	10335	PASS
75	95	30	80	50.8	29584	PASS
95	95	100	100	100.0	58189	PASS
96	95	5	9	6.7	3911	PASS
173	174	0.00	2	1.7	766	PASS
174	95	50	120	77.7	45197	PASS
175	174	5	9	8.0	3614	PASS
176	174	95	105	96.6	43667	PASS
177	176	5	10	6.9	3006	PASS

m/z	Abundance
40.00	459.0
43.90	277.0
51.00	102.0

Instrument :
MSVOA_V
ClientSampleId :
BFB292

m/z	Abundance
36.00	433.0
37.00	2509.0
38.00	2272.0
39.00	1151.0
40.00	326.0
44.00	462.0
45.10	537.0
47.00	819.0
48.00	352.0
49.00	2243.0
50.00	10177.0
51.00	3090.0
51.90	138.0
54.90	201.0
56.00	1046.0
57.00	1685.0
58.00	105.0
60.00	467.0
61.00	2667.0
62.00	2768.0
63.10	2531.0
64.00	445.0
65.00	525.0
66.80	115.0
68.00	5871.0
69.00	5891.0
70.00	486.0
72.10	328.0
73.00	2676.0
74.10	10198.0
75.00	28464.0
76.00	2432.0
77.00	297.0
77.80	200.0
78.90	2416.0
79.90	730.0
80.90	2553.0
81.80	507.0
87.00	2378.0
87.90	2795.0
90.80	286.0
92.00	1694.0
93.00	2898.0
94.00	7120.0
95.00	55664.0
96.10	3747.0
97.00	132.0
104.00	278.0
104.90	105.0
106.00	317.0
115.90	282.0
116.80	500.0
117.80	232.0
118.80	400.0
127.90	291.0
128.80	132.0
129.80	232.0
135.10	177.0
136.80	114.0
140.90	611.0
142.90	883.0
144.90	257.0
145.90	111.0
147.60	103.0
148.20	112.0
150.00	103.0
154.90	195.0
156.80	104.0
158.90	101.0
171.00	168.0
172.00	711.0
173.10	658.0
174.00	42216.0
175.10	3343.0
176.00	40464.0
177.00	2718.0

Instrument :
MSVOA_V
ClientSampleId :
BFB292

m/z	Abundance
36.00	497.0
37.00	2465.0
38.00	2503.0
39.00	1260.0
40.10	376.0
44.00	514.0
45.00	581.0
47.10	788.0
48.00	376.0
49.00	2389.0
50.00	10623.0
51.00	3206.0
51.90	194.0
54.90	236.0
56.00	1086.0
57.00	1726.0
59.90	573.0
61.00	2899.0
62.00	2978.0
63.10	2813.0
64.00	427.0
65.00	554.0
66.80	129.0
68.00	6264.0
69.00	6685.0
70.00	589.0
72.10	376.0
73.00	2953.0
74.10	11010.0
75.00	30520.0
76.10	2774.0
77.00	301.0
79.00	2490.0
79.90	786.0
80.90	2692.0
81.90	547.0
85.80	125.0
87.00	2560.0
87.90	2852.0
91.00	370.0
92.00	1953.0
93.00	3244.0
94.00	7542.0
95.00	60136.0
96.10	4006.0
97.00	124.0
103.90	284.0
104.90	117.0
106.00	305.0
110.90	106.0
115.90	285.0
116.80	466.0
117.80	259.0
118.90	356.0
127.90	307.0
128.70	120.0
129.80	255.0
135.00	199.0
137.00	126.0
140.90	666.0
142.00	116.0
142.90	872.0
144.90	277.0
145.90	117.0
147.70	164.0
148.10	121.0
154.90	215.0
158.90	132.0
170.90	192.0
172.00	777.0
173.00	779.0
174.00	46648.0
175.00	3750.0
176.00	45040.0
177.00	3160.0

Instrument :
MSVOA_V
ClientSampleId :
BFB292

m/z	Abundance
36.00	485.0
37.00	2341.0
38.00	2541.0
39.00	1182.0
40.00	371.0
44.00	629.0
45.00	540.0
45.80	119.0
47.10	692.0
48.00	374.0
49.00	2298.0
50.00	10204.0
51.00	2987.0
51.90	248.0
54.90	223.0
56.00	1010.0
57.00	1605.0
59.90	578.0
61.00	2955.0
62.00	2923.0
63.00	2592.0
64.00	395.0
65.00	527.0
66.80	111.0
67.20	114.0
68.00	6024.0
69.00	6634.0
70.00	581.0
72.00	366.0
73.00	2723.0
74.10	10386.0
75.00	29768.0
76.10	2861.0
77.00	360.0
78.10	265.0
79.00	2470.0
79.90	771.0
80.90	2520.0
81.90	595.0
85.90	112.0
87.00	2530.0
87.90	2711.0
91.00	453.0
92.00	1962.0
93.00	3163.0
94.10	7336.0
95.00	58768.0
96.10	3979.0
103.90	276.0
104.90	110.0
105.90	291.0
110.80	106.0
115.80	262.0
116.80	458.0
117.80	227.0
119.00	357.0
127.80	281.0
128.60	136.0
129.80	245.0
135.00	197.0
137.00	175.0
140.90	729.0
142.00	115.0
142.90	880.0
144.90	273.0
147.70	206.0
154.90	171.0
158.90	146.0
161.00	104.0
170.90	186.0
172.00	799.0
173.00	861.0
174.00	46728.0
175.00	3750.0
176.00	45496.0
177.00	3139.0

Instrument :
MSVOA_V
ClientSampleId :
BFB292