

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040221\
 Data File : VV020888.D
 Acq On : 02 Apr 2021 19:27
 Operator : SY/MD
 Sample : M1896-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3G6

Quant Time: Apr 03 00:40:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 03 00:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	622607	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	600079	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	292354	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	153305	34.68	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 69.36%			
7) Chloroethane-d5	1.571	69	134725	40.31	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 80.62%			
11) 1,1-Dichloroethene-d2	2.108	63	251567	30.17	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 60.34%			
21) 2-Butanone-d5	3.905	46	253853	88.27	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 88.27%			
24) Chloroform-d	4.352	84	421952	48.03	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 96.06%			
26) 1,2-Dichloroethane-d4	5.037	65	290329	50.17	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 100.34%			
32) Benzene-d6	5.053	84	769902	48.13	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 96.26%			
36) 1,2-Dichloropropane-d6	6.072	67	231508	47.14	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 94.28%			
41) Toluene-d8	7.320	98	741735	49.82	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 99.64%			
43) trans-1,3-Dichloroprop...	7.625	79	122042	46.71	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 93.42%			
47) 2-Hexanone-d5	8.092	63	176370	85.66	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 85.66%			
57) 1,1,2,2-Tetrachloroeth...	10.220	84	309799	43.31	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 86.62%			
64) 1,2-Dichlorobenzene-d4	11.632	152	314658	55.46	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 110.92%			
Target Compounds					Qvalue	
13) Acetone	2.198	43	14760	7.20	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040221\
Data File : VV020888.D
Acq On : 02 Apr 2021 19:27
Operator : SY/MD
Sample : M1896-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3G6

Quant Time: Apr 03 00:40:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 03 00:36:16 2021
Response via : Initial Calibration

