

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040221\
 Data File : VV020889.D
 Acq On : 02 Apr 2021 19:51
 Operator : SY/MD
 Sample : M1896-02
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3J0

Quant Time: Apr 03 01:59:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 03 00:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	620636	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	594368	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	290556	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	153576	34.85	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.70%
7) Chloroethane-d5	1.571	69	132045	39.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.28%
11) 1,1-Dichloroethene-d2	2.108	63	250842	30.17	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.34%
21) 2-Butanone-d5	3.899	46	248897	86.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.83%
24) Chloroform-d	4.352	84	430049	49.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.20%
26) 1,2-Dichloroethane-d4	5.037	65	293868	50.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.88%
32) Benzene-d6	5.053	84	794782	50.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.34%
36) 1,2-Dichloropropane-d6	6.072	67	234185	48.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.30%
41) Toluene-d8	7.320	98	754916	51.19	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	102.38%
43) trans-1,3-Dichloroprop...	7.625	79	127514	49.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.54%
47) 2-Hexanone-d5	8.095	63	174543	85.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.59%
57) 1,1,2,2-Tetrachloroeth...	10.223	84	315353	44.51	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.02%
64) 1,2-Dichlorobenzene-d4	11.631	152	325387	57.70	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.40%
Target Compounds						
13) Acetone	2.198	43	15565	7.61	ug/L	Qvalue 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040221\
Data File : VV020889.D
Acq On : 02 Apr 2021 19:51
Operator : SY/MD
Sample : M1896-02
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3J0

Quant Time: Apr 03 01:59:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 03 00:36:16 2021
Response via : Initial Calibration

