

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040221\
 Data File : VW020876.D
 Acq On : 02 Apr 2021 14:40
 Operator : SY/MD
 Sample : M1879-19
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1F0

Quant Time: Apr 03 00:37:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 03 00:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	700441	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	674744	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	312657	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	156757	31.52	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.04%
7) Chloroethane-d5	1.571	69	139995	37.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.48%
11) 1,1-Dichloroethene-d2	2.108	63	270428	28.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	57.64%#
21) 2-Butanone-d5	3.902	46	245767	75.97	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	75.97%
24) Chloroform-d	4.352	84	435195	44.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.06%
26) 1,2-Dichloroethane-d4	5.037	65	298154	45.79	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.58%
32) Benzene-d6	5.053	84	794270	44.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.32%
36) 1,2-Dichloropropane-d6	6.072	67	230596	41.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.52%
41) Toluene-d8	7.320	98	724208	43.26	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.52%
43) trans-1,3-Dichloroprop...	7.625	79	113355	38.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.16%
47) 2-Hexanone-d5	8.092	63	170559	73.68	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.68%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	305975	38.04	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	76.08%
64) 1,2-Dichlorobenzene-d4	11.632	152	306114	50.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.90%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040221\
Data File : VV020876.D
Acq On : 02 Apr 2021 14:40
Operator : SY/MD
Sample : M1879-19
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1F0

Quant Time: Apr 03 00:37:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 03 00:36:16 2021
Response via : Initial Calibration

