

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040221\
 Data File : VW020886.D
 Acq On : 02 Apr 2021 18:40
 Operator : SY/MD
 Sample : M1879-17
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1C3

Quant Time: Apr 03 00:40:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 03 00:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	610588	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	587230	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	283070	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	151625	34.98	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.96%
7) Chloroethane-d5	1.571	69	128759	39.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.58%
11) 1,1-Dichloroethene-d2	2.108	63	248848	30.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.86%
21) 2-Butanone-d5	3.902	46	245379	87.01	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	87.01%
24) Chloroform-d	4.352	84	419007	48.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.26%
26) 1,2-Dichloroethane-d4	5.037	65	294021	51.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.60%
32) Benzene-d6	5.053	84	770744	49.24	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.48%
36) 1,2-Dichloropropane-d6	6.072	67	232702	48.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.84%
41) Toluene-d8	7.320	98	730577	50.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.28%
43) trans-1,3-Dichloroprop...	7.625	79	125288	49.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.00%
47) 2-Hexanone-d5	8.095	63	171401	85.07	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	85.07%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	319868	45.70	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.40%
64) 1,2-Dichlorobenzene-d4	11.632	152	317035	57.71	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.42%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040221\
Data File : VV020886.D
Acq On : 02 Apr 2021 18:40
Operator : SY/MD
Sample : M1879-17
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1C3

Quant Time: Apr 03 00:40:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 03 00:36:16 2021
Response via : Initial Calibration

