

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040221\
 Data File : VW020887.D
 Acq On : 02 Apr 2021 19:04
 Operator : SY/MD
 Sample : M1879-18
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1E5

Quant Time: Apr 03 00:40:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 03 00:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	613810	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	592729	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	289336	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	153079	35.13	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.26%
7) Chloroethane-d5	1.571	69	131121	39.80	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.60%
11) 1,1-Dichloroethene-d2	2.108	63	253360	30.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	61.64%
21) 2-Butanone-d5	3.899	46	244996	86.42	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.42%
24) Chloroform-d	4.352	84	427688	49.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.76%
26) 1,2-Dichloroethane-d4	5.037	65	296349	51.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.88%
32) Benzene-d6	5.053	84	778185	49.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.52%
36) 1,2-Dichloropropane-d6	6.072	67	230311	47.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.96%
41) Toluene-d8	7.320	98	745113	50.67	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.34%
43) trans-1,3-Dichloroprop...	7.625	79	126006	48.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.64%
47) 2-Hexanone-d5	8.091	63	178469	87.76	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.76%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	321170	45.46	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.92%
64) 1,2-Dichlorobenzene-d4	11.628	152	317285	56.50	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	113.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040221\
Data File : VV020887.D
Acq On : 02 Apr 2021 19:04
Operator : SY/MD
Sample : M1879-18
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1E5

Quant Time: Apr 03 00:40:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 03 00:36:16 2021
Response via : Initial Calibration

