

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040224\
 Data File : VV034929.D
 Acq On : 02 Apr 2024 13:34
 Operator : SY/MD
 Sample : P1925-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 FB(20240327)

Quant Time: Apr 03 05:19:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 03 05:08:19 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.535	114	169577	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.786	117	170777	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.185	152	73453	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.278	65	74321	5.318	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.400%	
7) Chloroethane-d5		1.532	69	62070	5.387	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.800%	
11) 1,1-Dichloroethene-d2		2.060	65	33112	5.217	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	104.400%	
20) 2-Butanone-d5		3.815	46	76112	34.807	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	69.620%	
24) Chloroform-d		4.249	84	120345	5.322	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.400%	
26) 1,2-Dichloroethane-d4		4.947	65	58045	5.684	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.600%	
32) Benzene-d6		4.960	84	248558	5.564	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.200%	
36) 1,2-Dichloropropane-d6		5.989	67	74520	5.684	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.600%	
41) Toluene-d8		7.243	98	207479	5.113	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.200%	
43) trans-1,3-Dichloroprop...		7.561	79	24138	5.182	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.600%	
46) 2-Hexanone-d5		8.030	63	84692	52.980	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.960%	
56) 1,1,2,2-Tetrachloroeth...		10.153	84	50870	5.561	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	111.200%	
66) 1,2-Dichlorobenzene-d4		11.561	152	69857	6.038	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.800%#	
Target Compounds							
16) Methylene chloride		2.449	84	9021	0.639	ug/L	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040224\
Data File : VV034929.D
Acq On : 02 Apr 2024 13:34
Operator : SY/MD
Sample : P1925-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
FB(20240327)

Quant Time: Apr 03 05:19:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 03 05:08:19 2024
Response via : Initial Calibration

