

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040419\
 Data File : VV010108.D
 Acq On : 04 Apr 2019 14:38
 Operator : SY/MD
 Sample : VV0404SBL01
 Misc : 5.00G/10mL/MSVOA V/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK83

Quant Time: Apr 04 16:19:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 06:01:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	286474	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	273870	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	127419	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	80469	20.98	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	83.92%
7) Chloroethane-d5	1.57	69	95239	19.99	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.96%
10) 1,1-Dichloroethene-d2	2.13	63	158392	16.37	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.48%
20) 2-Butanone-d5	3.94	46	46927	43.00	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	86.00%
24) Chloroform-d	4.40	84	163503	22.65	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.60%
26) 1,2-Dichloroethane-d4	5.08	65	96501	23.90	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.60%
29) Benzene-d6	5.10	84	336853	23.72	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.88%
33) 1,2-Dichloropropane-d6	6.12	67	98423	23.48	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.92%
37) Toluene-d8	7.36	98	313808	22.93	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.72%
38) trans-1,3-Dichloropropene-	7.66	79	37673	21.03	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	84.12%
39) 2-Hexanone-d5	8.13	63	37204	44.26	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	88.52%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	91889	22.23	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.92%
61) 1,2-Dichlorobenzene-d4	11.67	152	118437	23.86	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.44%


Target Compounds					Ovalue
16) Methylene chloride	2.54	84	13621	2.706 ug/L	91




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV040419\
Data File : VV010108.D
Acq On : 04 Apr 2019 14:38
Operator : SY/MD
Sample : VV0404SBL01
Misc : 5.00G/10mL/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK83

Quant Time: Apr 04 16:19:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 06:01:11 2019
Response via : Initial Calibration





#16
Methvlene chloride
Concen: 2.706 ug/L
RT: 2.54 min Scan# 450
Delta R.T. -0.00 min
Lab File: VV010108.D
Acq: 04 Apr 2019 14:38

Instrument :
MSVOA_V
ClientSampleId :
VBLK83

Tot Ion: 84 Resp: 13621
Ion Ratio Lower Upper
84 100
49 109.8 84.3 156.5
86 58.4 45.4 84.2

