

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040424\
 Data File : VW034968.D
 Acq On : 04 Apr 2024 11:11
 Operator : SY/MD
 Sample : PB159974TB 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VLEB974

Quant Time: Apr 04 21:34:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 21:33:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	398220	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	398610	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	169759	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	158560	41.229	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.460%
7) Chloroethane-d5	1.535	69	145683	43.613	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.220%
11) 1,1-Dichloroethene-d2	2.059	65	74435	44.154	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	88.300%
21) 2-Butanone-d5	3.796	46	181795	79.704	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	79.700%
24) Chloroform-d	4.249	84	288288	44.585	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.160%
26) 1,2-Dichloroethane-d4	4.940	65	185323	45.143	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.280%
32) Benzene-d6	4.960	84	583966	43.614	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.220%
36) 1,2-Dichloropropane-d6	5.985	67	188917	44.432	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.860%
41) Toluene-d8	7.239	98	508735	42.189	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.380%
43) trans-1,3-Dichloroprop...	7.551	79	79090	38.263	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.520%
47) 2-Hexanone-d5	8.027	63	105181	65.430	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	65.430%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	238407	40.718	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.440%
66) 1,2-Dichlorobenzene-d4	11.558	152	180597	47.383	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.760%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040424\
Data File : VV034968.D
Acq On : 04 Apr 2024 11:11
Operator : SY/MD
Sample : PB159974TB 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VLEB974

Quant Time: Apr 04 21:34:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 21:33:57 2024
Response via : Initial Calibration

