

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040424\
 Data File : VW034970.D
 Acq On : 04 Apr 2024 11:59
 Operator : SY/MD
 Sample : P1893-02 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E07D1

Quant Time: Apr 04 21:35:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 21:33:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	407649	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	399927	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	172275	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	152655	38.776	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	77.560%
7) Chloroethane-d5	1.532	69	159427	46.624	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.240%
11) 1,1-Dichloroethene-d2	2.056	65	71260	41.292	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.580%
21) 2-Butanone-d5	3.796	46	198480	85.006	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.010%
24) Chloroform-d	4.246	84	291757	44.078	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.160%
26) 1,2-Dichloroethane-d4	4.937	65	189225	45.027	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.060%
32) Benzene-d6	4.957	84	586993	43.696	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.400%
36) 1,2-Dichloropropane-d6	5.985	67	192706	45.174	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.340%
41) Toluene-d8	7.239	98	508561	42.036	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.080%
43) trans-1,3-Dichloroprop...	7.551	79	81074	39.094	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.180%
47) 2-Hexanone-d5	8.027	63	120234	74.547	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	74.550%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	250277	42.604	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.200%
66) 1,2-Dichlorobenzene-d4	11.558	152	185022	47.835	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040424\
Data File : VV034970.D
Acq On : 04 Apr 2024 11:59
Operator : SY/MD
Sample : P1893-02 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E07D1

Quant Time: Apr 04 21:35:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 21:33:57 2024
Response via : Initial Calibration

