

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040424\
 Data File : VW034981.D
 Acq On : 04 Apr 2024 16:21
 Operator : SY/MD
 Sample : P1987-05
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MCOR66

Quant Time: Apr 04 21:37:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 21:33:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	369468	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	366904	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	152723	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	135138	37.873	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.740%
7) Chloroethane-d5	1.532	69	129803	41.883	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	83.760%
11) 1,1-Dichloroethene-d2	2.060	65	62156	39.739	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.480%
21) 2-Butanone-d5	3.796	46	190339	89.944	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.940%
24) Chloroform-d	4.246	84	262107	43.690	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.380%
26) 1,2-Dichloroethane-d4	4.941	65	171706	45.081	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.160%
32) Benzene-d6	4.960	84	517959	42.027	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.060%
36) 1,2-Dichloropropane-d6	5.986	67	175125	44.747	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	89.500%
41) Toluene-d8	7.243	98	440304	39.670	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.340%#
43) trans-1,3-Dichloroprop...	7.555	79	70658	37.138	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.280%
47) 2-Hexanone-d5	8.027	63	105053	70.997	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	71.000%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	233673	43.358	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.720%
66) 1,2-Dichlorobenzene-d4	11.558	152	163015	47.541	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.080%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040424\
Data File : VV034981.D
Acq On : 04 Apr 2024 16:21
Operator : SY/MD
Sample : P1987-05
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MC0R66

Quant Time: Apr 04 21:37:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 21:33:57 2024
Response via : Initial Calibration

