

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW040622\  
 Data File : VW025362.D  
 Acq On : 06 Apr 2022 14:38  
 Operator : SY/MD  
 Sample : N2262-21  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 9 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 EW6R2

Quant Time: Apr 07 01:26:47 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR032622WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Apr 07 01:24:47 2022  
 Response via : Initial Calibration

| Compound                      |        | R.T.   | QIon     | Response | Conc   | Units     | Dev(Min) |
|-------------------------------|--------|--------|----------|----------|--------|-----------|----------|
| -----                         |        |        |          |          |        |           |          |
| Internal Standards            |        |        |          |          |        |           |          |
| 1) 1,4-Difluorobenzene        |        | 5.619  | 114      | 119577   | 5.000  | ug/L      | 0.00     |
| 28) Chlorobenzene-d5          |        | 8.853  | 117      | 117540   | 5.000  | ug/L      | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    |        | 11.249 | 152      | 48516    | 5.000  | ug/L      | 0.00     |
|                               |        |        |          |          |        |           |          |
| System Monitoring Compounds   |        |        |          |          |        |           |          |
| 4) Vinyl Chloride-d3          |        | 1.307  | 65       | 66173    | 5.924  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 40 - 130 | Recovery | =      | 118.400%  |          |
| 7) Chloroethane-d5            |        | 1.568  | 69       | 66906    | 7.137  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 65 - 130 | Recovery | =      | 142.800%# |          |
| 11) 1,1-Dichloroethene-d2     |        | 2.108  | 63       | 107143   | 4.428  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 60 - 125 | Recovery | =      | 88.600%   |          |
| 20) 2-Butanone-d5             |        | 3.876  | 46       | 98365    | 88.409 | ug/L      | -0.02    |
| Spiked Amount                 | 50.000 | Range  | 40 - 130 | Recovery | =      | 176.820%# |          |
| 24) Chloroform-d              |        | 4.349  | 84       | 112768   | 7.736  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 70 - 125 | Recovery | =      | 154.800%# |          |
| 26) 1,2-Dichloroethane-d4     |        | 5.030  | 65       | 51543    | 8.000  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 70 - 130 | Recovery | =      | 160.000%# |          |
| 32) Benzene-d6                |        | 5.050  | 84       | 216912   | 7.295  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 70 - 125 | Recovery | =      | 145.800%# |          |
| 36) 1,2-Dichloropropane-d6    |        | 6.069  | 67       | 64783    | 7.774  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 60 - 140 | Recovery | =      | 155.400%# |          |
| 41) Toluene-d8                |        | 7.317  | 98       | 181853   | 6.667  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 70 - 130 | Recovery | =      | 133.400%# |          |
| 43) trans-1,3-Dichloroprop... |        | 7.625  | 79       | 18722    | 6.886  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 55 - 130 | Recovery | =      | 137.800%# |          |
| 46) 2-Hexanone-d5             |        | 8.085  | 63       | 75597    | 76.991 | ug/L      | 0.00     |
| Spiked Amount                 | 50.000 | Range  | 45 - 130 | Recovery | =      | 153.980%# |          |
| 56) 1,1,2,2-Tetrachloroeth... |        | 10.214 | 84       | 41213    | 7.417  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 65 - 120 | Recovery | =      | 148.400%# |          |
| 66) 1,2-Dichlorobenzene-d4    |        | 11.622 | 152      | 66670    | 8.556  | ug/L      | 0.00     |
| Spiked Amount                 | 5.000  | Range  | 80 - 120 | Recovery | =      | 171.200%# |          |
|                               |        |        |          |          |        |           |          |
| Target Compounds              |        |        |          |          |        |           | Qvalue   |
| 13) Acetone                   |        | 2.166  | 43       | 4185     | 4.929  | ug/L      | 99       |
| 16) Methylene chloride        |        | 2.506  | 84       | 5294     | 0.560  | ug/L      | 92       |
| 25) Chloroform                |        | 4.375  | 83       | 2646     | 0.165  | ug/L      | 98       |
| -----                         |        |        |          |          |        |           |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV040622\  
Data File : VV025362.D  
Acq On : 06 Apr 2022 14:38  
Operator : SY/MD  
Sample : N2262-21  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 9 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
EW6R2

Quant Time: Apr 07 01:26:47 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR032622WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Apr 07 01:24:47 2022  
Response via : Initial Calibration

