

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040722\
 Data File : VW025393.D
 Acq On : 07 Apr 2022 18:25
 Operator : SY/MD
 Sample : N2291-05DL 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0B32DL

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/08/2022
 Supervised By : Mahesh Dadoda 04/08/2022

Quant Time: Apr 08 01:24:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 08 01:19:10 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	156115	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	145490	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	69072	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	77554	5.318	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	= 106.400%		
7) Chloroethane-d5	1.568	69	87934	7.185	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	= 143.600%#		
11) 1,1-Dichloroethene-d2	2.105	63	113663	3.598	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	= 72.000%		
20) 2-Butanone-d5	3.892	46	86576	59.601	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	= 119.200%		
24) Chloroform-d	4.349	84	107365	5.642	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 112.800%		
26) 1,2-Dichloroethane-d4	5.034	65	51319	6.101	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 122.000%		
32) Benzene-d6	5.050	84	235425	6.396	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	= 128.000%#		
36) 1,2-Dichloropropane-d6	6.069	67	65158	6.317	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	= 126.400%		
41) Toluene-d8	7.317	98	207852	6.157	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	= 123.200%		
43) trans-1,3-Dichloroprop...	7.625	79	19032	5.656	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	= 113.200%		
46) 2-Hexanone-d5	8.092	63	73649	60.597	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	= 121.200%		
56) 1,1,2,2-Tetrachloroeth...	10.217	84	39465	5.738	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	= 114.800%		
66) 1,2-Dichlorobenzene-d4	11.625	152	70854	6.387	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	= 127.800%#		
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.748	101	68754	2.476	ug/L	99
16) Methylene chloride	2.507	84	3004	0.243	ug/L	96
22) cis-1,2-Dichloroethene	3.909	96	4362m	0.381	ug/L	
34) Trichloroethene	5.918	95	19055	1.626	ug/L	96
47) Tetrachloroethene	7.976	164	83343	9.645	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040722\
Data File : VV025393.D
Acq On : 07 Apr 2022 18:25
Operator : SY/MD
Sample : N2291-05DL 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0B32DL

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/08/2022
Supervised By : Mahesh Dadoda 04/08/2022

Quant Time: Apr 08 01:24:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032622WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 08 01:19:10 2022
Response via : Initial Calibration

