

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040822\
 Data File : VW025422.D
 Acq On : 08 Apr 2022 20:17
 Operator : SY/MD
 Sample : N2319-02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2G9

Quant Time: Apr 09 06:26:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 09 06:16:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	130627	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	124883	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56939	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	67424	5.552	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.000%
7) Chloroethane-d5	1.571	69	60467	4.680	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	93.600%
11) 1,1-Dichloroethene-d2	2.111	63	116167	4.775	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.400%
20) 2-Butanone-d5	3.895	46	73103	49.486	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.980%
24) Chloroform-d	4.352	84	88808	4.839	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.800%
26) 1,2-Dichloroethane-d4	5.037	65	40589	5.133	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.600%
32) Benzene-d6	5.053	84	188485	5.216	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.400%
36) 1,2-Dichloropropane-d6	6.072	67	54280	5.386	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.800%
41) Toluene-d8	7.317	98	168136	5.014	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.200%
43) trans-1,3-Dichloroprop...	7.625	79	14984	4.424	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	88.400%
46) 2-Hexanone-d5	8.088	63	57944	48.743	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.480%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32309	5.075	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	101.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	57274	6.000	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	120.000%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.757	101	6292	0.309	ug/L	94
10) 1,1,2-Trichloro-1,2,2-...	2.117	101	3314	0.290	ug/L #	81
12) 1,1-Dichloroethene	2.121	96	28055	2.476	ug/L #	53
18) trans-1,2-Dichloroethene	2.767	96	3107	0.327	ug/L	91
19) 1,1-Dichloroethane	3.191	63	2507	0.156	ug/L #	86
22) cis-1,2-Dichloroethene	3.915	96	28693	2.871	ug/L	91
25) Chloroform	4.381	83	16701	0.984	ug/L	85
31) Carbon tetrachloride	4.831	117	8852	0.711	ug/L	86
34) Trichloroethene	5.911	95	714123	72.907	ug/L	98
38) Bromodichloromethane	6.526	83	869	0.082	ug/L #	75
47) Tetrachloroethene	7.976	164	56758	7.402	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040822\
Data File : VV025422.D
Acq On : 08 Apr 2022 20:17
Operator : SY/MD
Sample : N2319-02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2G9

Quant Time: Apr 09 06:26:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 09 06:16:16 2022
Response via : Initial Calibration

