

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040822\
 Data File : VW025423.D
 Acq On : 08 Apr 2022 20:41
 Operator : SY/MD
 Sample : N2319-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2H0

Quant Time: Apr 09 06:26:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 09 06:16:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	126490	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	116879	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53337	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	66152	5.625	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	112.600%
7) Chloroethane-d5	1.568	69	60398	4.828	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	96.600%
11) 1,1-Dichloroethene-d2	2.111	63	113619	4.823	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	96.400%
20) 2-Butanone-d5	3.892	46	71024	49.651	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.300%
24) Chloroform-d	4.352	84	88090	4.957	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.200%
26) 1,2-Dichloroethane-d4	5.034	65	40913	5.343	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	106.800%
32) Benzene-d6	5.053	84	188766	5.582	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	111.600%
36) 1,2-Dichloropropane-d6	6.072	67	53580	5.681	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	113.600%
41) Toluene-d8	7.317	98	166032	5.291	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.800%
43) trans-1,3-Dichloroprop...	7.629	79	14171	4.470	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.400%
46) 2-Hexanone-d5	8.088	63	55562	49.940	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.880%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31276	5.249	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	55148	6.167	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	123.400%#
Target Compounds						Qvalue
9) Trichlorofluoromethane	1.754	101	6404	0.325	ug/L	99
10) 1,1,2-Trichloro-1,2,2-...	2.118	101	3309	0.299	ug/L #	71
12) 1,1-Dichloroethene	2.121	96	25435	2.318	ug/L #	65
18) trans-1,2-Dichloroethene	2.770	96	3067	0.334	ug/L	91
19) 1,1-Dichloroethane	3.195	63	2375	0.153	ug/L	96
22) cis-1,2-Dichloroethene	3.912	96	27964	2.889	ug/L	96
25) Chloroform	4.378	83	16785	1.021	ug/L	86
31) Carbon tetrachloride	4.828	117	8017	0.688	ug/L	97
34) Trichloroethene	5.912	95	706045	77.018	ug/L	99
38) Bromodichloromethane	6.519	83	818	0.082	ug/L #	91
47) Tetrachloroethene	7.976	164	54902	7.650	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040822\
Data File : VV025423.D
Acq On : 08 Apr 2022 20:41
Operator : SY/MD
Sample : N2319-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2H0

Quant Time: Apr 09 06:26:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 09 06:16:16 2022
Response via : Initial Calibration

