

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040822\
 Data File : VV025429.D
 Acq On : 08 Apr 2022 23:02
 Operator : SY/MD
 Sample : N2319-11
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2H9

Quant Time: Apr 09 06:28:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 09 06:16:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	124784	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	120884	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53729	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	64539	5.563	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	111.200%	
7) Chloroethane-d5	1.571	69	57859	4.688	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	93.800%	
11) 1,1-Dichloroethene-d2	2.111	63	95290	4.100	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	82.000%	
20) 2-Butanone-d5	3.892	46	75574	53.554	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	107.100%	
24) Chloroform-d	4.352	84	84562	4.824	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.400%	
26) 1,2-Dichloroethane-d4	5.037	65	41124	5.444	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.800%	
32) Benzene-d6	5.053	84	177560	5.077	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	101.600%	
36) 1,2-Dichloropropane-d6	6.072	67	51584	5.288	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	105.800%	
41) Toluene-d8	7.316	98	157978	4.867	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	97.400%	
43) trans-1,3-Dichloroprop...	7.628	79	14142	4.313	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.200%	
46) 2-Hexanone-d5	8.088	63	62023	53.901	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	107.800%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33660	5.462	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	109.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	53671	5.958	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	119.200%	
Target Compounds						
9) Trichlorofluoromethane	1.754	101	893	0.046	ug/L	95
17) Methyl tert-butyl Ether	2.770	73	2538	0.152	ug/L #	89
22) cis-1,2-Dichloroethene	3.918	96	8441	0.884	ug/L	97
25) Chloroform	4.381	83	13155	0.811	ug/L	89
34) Trichloroethene	5.911	95	465076	49.051	ug/L	99
38) Bromodichloromethane	6.522	83	618	0.060	ug/L #	91
45) 1,1,2-Trichloroethane	7.853	97	507	0.092	ug/L	90
47) Tetrachloroethene	7.976	164	307144	41.378	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040822\
Data File : VV025429.D
Acq On : 08 Apr 2022 23:02
Operator : SY/MD
Sample : N2319-11
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2H9

Quant Time: Apr 09 06:28:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 09 06:16:16 2022
Response via : Initial Calibration

