

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041122\
 Data File : VW025445.D
 Acq On : 11 Apr 2022 13:15
 Operator : SY/MD
 Sample : N2319-06DL 10X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2H3DL

Quant Time: Apr 12 07:05:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 12 07:03:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	151764	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	151487	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	72419	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	69923	4.956	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	99.200%	
7) Chloroethane-d5	1.571	69	63856	4.254	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	85.000%	
11) 1,1-Dichloroethene-d2	2.108	63	116037	4.105	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	82.200%	
20) 2-Butanone-d5	3.902	46	90136	52.518	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	105.040%	
24) Chloroform-d	4.349	84	105885	4.966	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.400%	
26) 1,2-Dichloroethane-d4	5.034	65	49769	5.417	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.400%	
32) Benzene-d6	5.050	84	212042	4.838	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.800%	
36) 1,2-Dichloropropane-d6	6.069	67	61655	5.044	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	100.800%	
41) Toluene-d8	7.317	98	184179	4.528	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	90.600%	
43) trans-1,3-Dichloroprop...	7.625	79	18038	4.390	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.800%	
46) 2-Hexanone-d5	8.091	63	70757	49.069	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	98.140%	
56) 1,1,2,2-Tetrachloroeth...	10.214	84	38043	4.926	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	62259	5.128	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	102.600%	
Target Compounds						
25) Chloroform	4.378	83	3834	0.194	ug/L	97
34) Trichloroethene	5.915	95	73724	6.205	ug/L	98
47) Tetrachloroethene	7.976	164	15939	1.714	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041122\
Data File : VV025445.D
Acq On : 11 Apr 2022 13:15
Operator : SY/MD
Sample : N2319-06DL 10X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2H3DL

Quant Time: Apr 12 07:05:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 12 07:03:45 2022
Response via : Initial Calibration

