

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041122\
 Data File : VW025450.D
 Acq On : 11 Apr 2022 15:12
 Operator : SY/MD
 Sample : N2319-07 100X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2H4

Quant Time: Apr 12 07:06:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 12 07:03:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	129751	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	129687	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	60665	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	61095	5.065	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	101.200%	
7) Chloroethane-d5	1.568	69	56236	4.382	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	87.600%	
11) 1,1-Dichloroethene-d2	2.108	63	93491	3.869	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	77.400%	
20) 2-Butanone-d5	3.902	46	76564	52.178	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	104.360%	
24) Chloroform-d	4.352	84	90535	4.967	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.400%	
26) 1,2-Dichloroethane-d4	5.037	65	42615	5.426	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	108.600%	
32) Benzene-d6	5.050	84	180297	4.805	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.000%	
36) 1,2-Dichloropropane-d6	6.072	67	51903	4.960	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.200%	
41) Toluene-d8	7.317	98	152603	4.382	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.600%	
43) trans-1,3-Dichloroprop...	7.625	79	15419	4.384	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	87.600%	
46) 2-Hexanone-d5	8.092	63	58135	47.093	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	94.180%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32530	4.920	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	98.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	53107	5.221	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	104.400%	
Target Compounds						Qvalue
34) Trichloroethene	5.921	95	2993	0.294	ug/L	90
47) Tetrachloroethene	7.979	164	5256	0.660	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041122\
Data File : VV025450.D
Acq On : 11 Apr 2022 15:12
Operator : SY/MD
Sample : N2319-07 100X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2H4

Quant Time: Apr 12 07:06:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 12 07:03:45 2022
Response via : Initial Calibration

