

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041122\
 Data File : VV025453.D
 Acq On : 11 Apr 2022 16:23
 Operator : SY/MD
 Sample : N2319-12
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2J0

Quant Time: Apr 12 07:07:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 12 07:03:45 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	130273	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	132411	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62174	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	59839	4.941	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.800%
7) Chloroethane-d5	1.571	69	56491	4.384	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%
11) 1,1-Dichloroethene-d2	2.111	63	91941	3.789	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	75.800%
20) 2-Butanone-d5	3.895	46	74465	50.545	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	101.080%
24) Chloroform-d	4.352	84	89134	4.870	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
26) 1,2-Dichloroethane-d4	5.034	65	41127	5.215	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.400%
32) Benzene-d6	5.053	84	173191	4.521	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.400%
36) 1,2-Dichloropropane-d6	6.069	67	51576	4.827	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	96.600%
41) Toluene-d8	7.317	98	145853	4.102	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.000%
43) trans-1,3-Dichloroprop...	7.625	79	15117	4.209	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.200%
46) 2-Hexanone-d5	8.091	63	54218	43.016	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.040%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31984	4.738	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	11.625	152	51044	4.897	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.000%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.130	85	639	0.054	ug/L	92
25) Chloroform	4.378	83	18655	1.102	ug/L	89
31) Carbon tetrachloride	4.831	117	38931	2.948	ug/L	98
34) Trichloroethene	5.921	95	12711	1.224	ug/L	97
38) Bromodichloromethane	6.516	83	1249	0.111	ug/L #	95
47) Tetrachloroethene	7.976	164	11605	1.427	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041122\
Data File : VV025453.D
Acq On : 11 Apr 2022 16:23
Operator : SY/MD
Sample : N2319-12
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2J0

Quant Time: Apr 12 07:07:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 12 07:03:45 2022
Response via : Initial Calibration

