

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041222\
 Data File : VW025468.D
 Acq On : 12 Apr 2022 11:47
 Operator : SY/MD
 Sample : N2319-10
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YB2H8

Quant Time: Apr 13 02:32:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 13 02:31:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	142284	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	143455	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	68882	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	62263	4.707	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.200%
7) Chloroethane-d5	1.568	69	61283	4.355	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.000%
11) 1,1-Dichloroethene-d2	2.108	63	102522	3.869	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.400%
20) 2-Butanone-d5	3.905	46	98375	61.137	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	122.280%
24) Chloroform-d	4.349	84	103481	5.177	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.600%
26) 1,2-Dichloroethane-d4	5.034	65	48915	5.679	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	113.600%
32) Benzene-d6	5.050	84	197750	4.764	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	95.200%
36) 1,2-Dichloropropane-d6	6.069	67	60356	5.214	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.200%
41) Toluene-d8	7.317	98	170393	4.424	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
43) trans-1,3-Dichloroprop...	7.625	79	17339	4.456	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	89.200%
46) 2-Hexanone-d5	8.088	63	74448	54.519	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	109.040%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	38807	5.306	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.200%
66) 1,2-Dichlorobenzene-d4	11.622	152	60034	5.198	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.000%
Target Compounds						
					Qvalue	
3) Chloromethane	1.240	50	1711	0.123	ug/L	96
13) Acetone	2.201	43	5252	4.809	ug/L	68
25) Chloroform	4.381	83	11106	0.601	ug/L	97
31) Carbon tetrachloride	4.828	117	764	0.053	ug/L	97
34) Trichloroethene	5.915	95	129694	11.527	ug/L	97
47) Tetrachloroethene	7.976	164	56594	6.425	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041222\
Data File : VV025468.D
Acq On : 12 Apr 2022 11:47
Operator : SY/MD
Sample : N2319-10
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YB2H8

Quant Time: Apr 13 02:32:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 13 02:31:13 2022
Response via : Initial Calibration

