

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041322\
 Data File : VV025502.D
 Acq On : 13 Apr 2022 19:07
 Operator : SY/MD
 Sample : N2356-19
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Apr 14 03:13:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 14 03:09:10 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	135633	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	130095	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	58682	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	64948	5.151	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	103.000%	
7) Chloroethane-d5	1.571	69	59026	4.400	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	88.000%	
11) 1,1-Dichloroethene-d2	2.111	63	100757	3.989	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	79.800%	
20) 2-Butanone-d5	3.896	46	85103	55.482	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	110.960%	
24) Chloroform-d	4.352	84	92699	4.865	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.200%	
26) 1,2-Dichloroethane-d4	5.034	65	42664	5.196	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	104.000%	
32) Benzene-d6	5.053	84	186539	4.956	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	99.200%	
36) 1,2-Dichloropropane-d6	6.072	67	51498	4.906	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.200%	
41) Toluene-d8	7.317	98	152251	4.359	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.200%	
43) trans-1,3-Dichloroprop...	7.625	79	14908	4.225	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.600%	
46) 2-Hexanone-d5	8.088	63	56149	45.341	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.680%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31116	4.692	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	93.800%	
66) 1,2-Dichlorobenzene-d4	11.622	152	48536	4.933	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	98.600%	
Target Compounds						
						Qvalue
5) Vinyl chloride	1.314	62	7806	0.519	ug/L	89
18) trans-1,2-Dichloroethene	2.767	96	27485	2.788	ug/L	93
22) cis-1,2-Dichloroethene	3.912	96	3873505	373.242	ug/L	94
25) Chloroform	4.381	83	18717	1.062	ug/L	88
34) Trichloroethene	5.915	95	21437009m	2100.875	ug/L	
38) Bromodichloromethane	6.519	83	2498	0.226	ug/L #	95
47) Tetrachloroethene	7.979	164	16315	2.042	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041322\
Data File : VV025502.D
Acq On : 13 Apr 2022 19:07
Operator : SY/MD
Sample : N2356-19
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Apr 14 03:13:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 14 03:09:10 2022
Response via : Initial Calibration

