

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041322\
 Data File : VW025506.D
 Acq On : 13 Apr 2022 21:04
 Operator : SY/MD
 Sample : VW0413WBL02
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK342

Quant Time: Apr 14 05:23:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 14 05:22:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	114962	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	114268	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	51834	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	65327	6.112	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	122.200%	
7) Chloroethane-d5	1.571	69	56086	4.933	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	98.600%	
11) 1,1-Dichloroethene-d2	2.111	63	96263	4.496	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	90.000%	
20) 2-Butanone-d5	3.905	46	66484	51.137	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.280%	
24) Chloroform-d	4.352	84	88555	5.483	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.600%	
26) 1,2-Dichloroethane-d4	5.037	65	39915	5.736	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	114.800%	
32) Benzene-d6	5.053	84	181118	5.478	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.600%	
36) 1,2-Dichloropropane-d6	6.072	67	51887	5.627	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	112.600%	
41) Toluene-d8	7.317	98	150710	4.912	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	98.200%	
43) trans-1,3-Dichloroprop...	7.629	79	14002	4.518	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	90.400%	
46) 2-Hexanone-d5	8.092	63	51821	47.642	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.280%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	31074	5.334	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.600%	
66) 1,2-Dichlorobenzene-d4	11.622	152	48460	5.576	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	111.600%	
Target Compounds						
6) Bromomethane	1.529	94	427	0.046	ug/L	99
34) Trichloroethene	5.924	95	3618	0.404	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041322\
Data File : VV025506.D
Acq On : 13 Apr 2022 21:04
Operator : SY/MD
Sample : VV0413WBL02
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK342

Quant Time: Apr 14 05:23:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 14 05:22:13 2022
Response via : Initial Calibration

