

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW041322\  
 Data File : VW025507.D  
 Acq On : 13 Apr 2022 21:51  
 Operator : SY/MD  
 Sample : N2357-07  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 24 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 YAZ04

Quant Time: Apr 14 05:23:50 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR040822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Thu Apr 14 05:22:13 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 5.619  | 114   | 116698   | 5.000    | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 8.854  | 117   | 119592   | 5.000    | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 54976    | 5.000    | ug/L   | 0.00     |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 4) Vinyl Chloride-d3          | 1.311  | 65    | 53535    | 4.934    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =      | 98.600%  |
| 7) Chloroethane-d5            | 1.571  | 69    | 49828    | 4.317    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =      | 86.400%  |
| 11) 1,1-Dichloroethene-d2     | 2.111  | 63    | 82535    | 3.797    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =      | 76.000%  |
| 20) 2-Butanone-d5             | 3.918  | 46    | 71403    | 54.104   | ug/L   | 0.01     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =      | 108.200% |
| 24) Chloroform-d              | 4.352  | 84    | 84206    | 5.136    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 102.800% |
| 26) 1,2-Dichloroethane-d4     | 5.037  | 65    | 40412    | 5.721    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 114.400% |
| 32) Benzene-d6                | 5.053  | 84    | 163098   | 4.713    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 94.200%  |
| 36) 1,2-Dichloropropane-d6    | 6.069  | 67    | 49776    | 5.158    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =      | 103.200% |
| 41) Toluene-d8                | 7.317  | 98    | 131038   | 4.081    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 81.600%  |
| 43) trans-1,3-Dichloroprop... | 7.629  | 79    | 11425    | 3.522    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =      | 70.400%  |
| 46) 2-Hexanone-d5             | 8.092  | 63    | 54775    | 48.116   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =      | 96.240%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 32775    | 5.376    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =      | 107.600% |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 47139    | 5.114    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =      | 102.200% |
| Target Compounds              |        |       |          |          |        |          |
|                               |        |       |          |          | Qvalue |          |
| 14) Carbon disulfide          | 2.298  | 76    | 4921     | 0.203    | ug/L   | 98       |
| 34) Trichloroethene           | 5.928  | 95    | 2355     | 0.251    | ug/L   | 95       |
| 53) m,p-Xylene                | 9.156  | 106   | 760      | 0.049    | ug/L   | 77       |
| -----                         |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV041322\  
Data File : VV025507.D  
Acq On : 13 Apr 2022 21:51  
Operator : SY/MD  
Sample : N2357-07  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 24 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
YAZ04

Quant Time: Apr 14 05:23:50 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR040822WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Thu Apr 14 05:22:13 2022  
Response via : Initial Calibration

