

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041322\
 Data File : VW025515.D
 Acq On : 14 Apr 2022 00:58
 Operator : SY/MD
 Sample : N2388-03
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFFE6

Quant Time: Apr 14 05:24:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 14 05:22:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	123682	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	124983	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	56352	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	62562	5.441	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	108.800%
7) Chloroethane-d5	1.571	69	55874	4.567	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	91.400%
11) 1,1-Dichloroethene-d2	2.111	63	92753	4.027	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	80.600%
20) 2-Butanone-d5	3.902	46	66388	47.463	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	94.920%
24) Chloroform-d	4.352	84	84690	4.874	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.400%
26) 1,2-Dichloroethane-d4	5.037	65	40144	5.362	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.200%
32) Benzene-d6	5.053	84	173499	4.798	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	96.000%
36) 1,2-Dichloropropane-d6	6.069	67	50407	4.998	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.000%
41) Toluene-d8	7.316	98	147511	4.396	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.000%
43) trans-1,3-Dichloroprop...	7.628	79	13151	3.880	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.600%
46) 2-Hexanone-d5	8.091	63	47068	39.563	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	79.120%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	28469	4.468	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.400%
66) 1,2-Dichlorobenzene-d4	11.625	152	47864	5.066	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.400%
Target Compounds						
					Qvalue	
13) Acetone	2.191	43	3898	4.106	ug/L	86
25) Chloroform	4.384	83	14933	0.929	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041322\
Data File : VV025515.D
Acq On : 14 Apr 2022 00:58
Operator : SY/MD
Sample : N2388-03
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFFE6

Quant Time: Apr 14 05:24:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 14 05:22:13 2022
Response via : Initial Calibration

