

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041322\
 Data File : VW025520.D
 Acq On : 14 Apr 2022 02:55
 Operator : SY/MD
 Sample : N2384-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6P4

Quant Time: Apr 14 05:25:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 14 05:22:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	117606	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	119326	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	50870	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	59471	5.439	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	108.800%	
7) Chloroethane-d5	1.571	69	55954	4.810	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	96.200%	
11) 1,1-Dichloroethene-d2	2.111	63	90622	4.137	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	82.800%	
20) 2-Butanone-d5	3.918	46	68306	51.358	ug/L	0.01
Spiked Amount	50.000	Range 40 - 130	Recovery	=	102.720%	
24) Chloroform-d	4.352	84	87338	5.286	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	105.800%	
26) 1,2-Dichloroethane-d4	5.037	65	40620	5.706	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	114.200%	
32) Benzene-d6	5.053	84	176668	5.117	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	102.400%	
36) 1,2-Dichloropropane-d6	6.072	67	50424	5.237	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	104.800%	
41) Toluene-d8	7.317	98	143225	4.470	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	89.400%	
43) trans-1,3-Dichloroprop...	7.625	79	14058	4.344	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	86.800%	
46) 2-Hexanone-d5	8.091	63	52774	46.462	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	92.920%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	32676	5.372	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	107.400%	
66) 1,2-Dichlorobenzene-d4	11.622	152	45564	5.342	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.800%	
Target Compounds						
2) Dichlorodifluoromethane	1.134	85	183588	17.281	ug/L	99
5) Vinyl chloride	1.314	62	1325	0.102	ug/L #	28
22) cis-1,2-Dichloroethene	3.931	96	1100	0.122	ug/L	96
25) Chloroform	4.381	83	1943	0.127	ug/L	90
34) Trichloroethene	5.928	95	2103	0.225	ug/L	88
53) m,p-Xylene	9.146	106	1156	0.074	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041322\
Data File : VV025520.D
Acq On : 14 Apr 2022 02:55
Operator : SY/MD
Sample : N2384-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6P4

Quant Time: Apr 14 05:25:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 14 05:22:13 2022
Response via : Initial Calibration

