

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041322\
 Data File : VW025525.D
 Acq On : 14 Apr 2022 04:52
 Operator : SY/MD
 Sample : N2373-05
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Apr 14 05:26:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 14 05:22:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	131781	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	133579	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	62604	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	58977	4.814	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.200%
7) Chloroethane-d5	1.571	69	52687	4.042	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	80.800%
11) 1,1-Dichloroethene-d2	2.114	63	87378	3.560	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.200%
20) 2-Butanone-d5	3.905	46	64166	43.055	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.120%
24) Chloroform-d	4.355	84	84633	4.571	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
26) 1,2-Dichloroethane-d4	5.037	65	40508	5.078	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
32) Benzene-d6	5.053	84	170474	4.411	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
36) 1,2-Dichloropropane-d6	6.072	67	49652	4.606	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.200%
41) Toluene-d8	7.317	98	142981	3.986	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.800%
43) trans-1,3-Dichloroprop...	7.629	79	13542	3.738	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	74.800%
46) 2-Hexanone-d5	8.092	63	49176	38.675	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	77.340%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	29750	4.369	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	47133	4.490	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.800%
Target Compounds						
16) Methylene chloride	2.513	84	4296	0.388	ug/L	93
22) cis-1,2-Dichloroethene	3.918	96	2334	0.231	ug/L	84
34) Trichloroethene	5.918	95	16892	1.612	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041322\
Data File : VV025525.D
Acq On : 14 Apr 2022 04:52
Operator : SY/MD
Sample : N2373-05
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Apr 14 05:26:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 14 05:22:13 2022
Response via : Initial Calibration

