

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041322\
 Data File : VW025513.D
 Acq On : 14 Apr 2022 00:11
 Operator : SY/MD
 Sample : N2357-15
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAZ65

Quant Time: Apr 14 05:24:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 14 05:22:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	125693	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	129114	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59332	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	54872	4.696	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	94.000%
7) Chloroethane-d5	1.571	69	51251	4.123	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	82.400%
11) 1,1-Dichloroethene-d2	2.111	63	85139	3.637	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	72.800%
20) 2-Butanone-d5	3.905	46	74967	52.739	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	105.480%
24) Chloroform-d	4.352	84	86191	4.881	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.600%
26) 1,2-Dichloroethane-d4	5.037	65	40790	5.361	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.200%
32) Benzene-d6	5.053	84	170669	4.569	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.400%
36) 1,2-Dichloropropane-d6	6.072	67	49743	4.775	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	95.400%
41) Toluene-d8	7.317	98	140769	4.060	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.200%
43) trans-1,3-Dichloroprop...	7.629	79	11398	3.255	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.000%
46) 2-Hexanone-d5	8.092	63	50562	41.140	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.280%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	29920	4.546	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	48851	4.911	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041322\
Data File : VV025513.D
Acq On : 14 Apr 2022 00:11
Operator : SY/MD
Sample : N2357-15
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAZ65

Quant Time: Apr 14 05:24:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 14 05:22:13 2022
Response via : Initial Calibration

