

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041422\
 Data File : VW025538.D
 Acq On : 14 Apr 2022 12:09
 Operator : SY/MD
 Sample : N2356-04DL 5X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDS2DL

Quant Time: Apr 15 07:21:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 15 07:05:00 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		5.619	114	125820	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.853	117	125231	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.249	152	57704	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.310	65	59791	5.111	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	102.200%	
7) Chloroethane-d5		1.574	69	54469	4.377	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	87.600%	
11) 1,1-Dichloroethene-d2		2.114	63	91373	3.899	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	78.000%	
20) 2-Butanone-d5		3.905	46	73211	51.452	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.900%	
24) Chloroform-d		4.355	84	91894	5.199	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.000%	
26) 1,2-Dichloroethane-d4		5.037	65	41968	5.510	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.200%	
32) Benzene-d6		5.053	84	178270	4.920	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.400%	
36) 1,2-Dichloropropane-d6		6.072	67	52073	5.153	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.000%	
41) Toluene-d8		7.317	98	149442	4.444	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.800%	
43) trans-1,3-Dichloroprop...		7.628	79	14597	4.298	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%	
46) 2-Hexanone-d5		8.091	63	51811	43.463	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	86.920%	
56) 1,1,2,2-Tetrachloroeth...		10.217	84	30483	4.775	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.400%	
66) 1,2-Dichlorobenzene-d4		11.622	152	48412	5.004	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.000%	
Target Compounds							
5) Vinyl chloride		1.317	62	5143	0.369	ug/L	# 82
18) trans-1,2-Dichloroethene		2.770	96	5080	0.556	ug/L	97
22) cis-1,2-Dichloroethene		3.918	96	50895	5.287	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041422\
Data File : VV025538.D
Acq On : 14 Apr 2022 12:09
Operator : SY/MD
Sample : N2356-04DL 5X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDS2DL

Quant Time: Apr 15 07:21:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 15 07:05:00 2022
Response via : Initial Calibration

