

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW041422\  
 Data File : VW025534.D  
 Acq On : 14 Apr 2022 10:36  
 Operator : SY/MD  
 Sample : N2357-17  
 Misc : 25mL/MSVOA\_V/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK001

Quant Time: Apr 15 07:20:51 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR040822WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Fri Apr 15 07:05:00 2022  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|-------|----------|
| -----                         |        |       |          |          |       |          |
| Internal Standards            |        |       |          |          |       |          |
| 1) 1,4-Difluorobenzene        | 5.616  | 114   | 180330   | 5.000    | ug/L  | 0.00     |
| 28) Chlorobenzene-d5          | 8.850  | 117   | 179027   | 5.000    | ug/L  | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.246 | 152   | 84657    | 5.000    | ug/L  | 0.00     |
|                               |        |       |          |          |       |          |
| System Monitoring Compounds   |        |       |          |          |       |          |
| 4) Vinyl Chloride-d3          | 1.310  | 65    | 72763    | 4.340    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =     | 86.800%  |
| 7) Chloroethane-d5            | 1.571  | 69    | 66969    | 3.755    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =     | 75.000%  |
| 11) 1,1-Dichloroethene-d2     | 2.111  | 63    | 112458   | 3.348    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =     | 67.000%  |
| 20) 2-Butanone-d5             | 3.895  | 46    | 83717    | 41.051   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =     | 82.100%  |
| 24) Chloroform-d              | 4.349  | 84    | 107696   | 4.251    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 85.000%  |
| 26) 1,2-Dichloroethane-d4     | 5.034  | 65    | 49456    | 4.531    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 90.600%  |
| 32) Benzene-d6                | 5.050  | 84    | 219353   | 4.235    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =     | 84.600%  |
| 36) 1,2-Dichloropropane-d6    | 6.069  | 67    | 63321    | 4.383    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =     | 87.600%  |
| 41) Toluene-d8                | 7.313  | 98    | 191987   | 3.994    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =     | 79.800%  |
| 43) trans-1,3-Dichloroprop... | 7.622  | 79    | 17970    | 3.701    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =     | 74.000%  |
| 46) 2-Hexanone-d5             | 8.088  | 63    | 63382    | 37.193   | ug/L  | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =     | 74.380%  |
| 56) 1,1,2,2-Tetrachloroeth... | 10.214 | 84    | 37438    | 4.102    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =     | 82.000%  |
| 66) 1,2-Dichlorobenzene-d4    | 11.622 | 152   | 61148    | 4.308    | ug/L  | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =     | 86.200%  |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV041422\  
Data File : VV025534.D  
Acq On : 14 Apr 2022 10:36  
Operator : SY/MD  
Sample : N2357-17  
Misc : 25mL/MSVOA\_V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VHBLK001

Quant Time: Apr 15 07:20:51 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR040822WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Fri Apr 15 07:05:00 2022  
Response via : Initial Calibration

