

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025562.D
 Acq On : 15 Apr 2022 12:59
 Operator : SY/MD
 Sample : N2388-05
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFFE8

Quant Time: Apr 16 02:42:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 02:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	130610	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	130204	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	59221	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	45077	3.712	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.200%
7) Chloroethane-d5	1.574	69	44251	3.425	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.600%
11) 1,1-Dichloroethene-d2	2.114	63	69829	2.871	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.400%#
20) 2-Butanone-d5	3.902	46	65497	44.343	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	88.680%
24) Chloroform-d	4.355	84	75044	4.090	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.800%
26) 1,2-Dichloroethane-d4	5.037	65	34915	4.416	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.400%
32) Benzene-d6	5.053	84	146308	3.884	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.600%
36) 1,2-Dichloropropane-d6	6.072	67	44532	4.239	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.800%
41) Toluene-d8	7.317	98	120178	3.438	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	68.800%#
43) trans-1,3-Dichloroprop...	7.628	79	11630	3.293	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	65.800%
46) 2-Hexanone-d5	8.091	63	42829	34.556	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	69.120%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	24673	3.717	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	74.400%
66) 1,2-Dichlorobenzene-d4	11.622	152	41545	4.184	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.600%
Target Compounds					Qvalue	
25) Chloroform	4.384	83	8843	0.521	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025562.D
Acq On : 15 Apr 2022 12:59
Operator : SY/MD
Sample : N2388-05
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFFE8

Quant Time: Apr 16 02:42:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 02:38:36 2022
Response via : Initial Calibration

