

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025565.D
 Acq On : 15 Apr 2022 14:09
 Operator : SY/MD
 Sample : N2356-10DL 100X
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 GBDS8DL

Quant Time: Apr 16 02:42:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 02:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	158880	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	155253	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	70533	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	54467	3.687	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	73.800%
7) Chloroethane-d5	1.574	69	54091	3.442	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	68.800%
11) 1,1-Dichloroethene-d2	2.114	63	83774	2.831	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	56.600%#
20) 2-Butanone-d5	3.899	46	92233	51.333	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	102.660%
24) Chloroform-d	4.352	84	96383	4.318	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
26) 1,2-Dichloroethane-d4	5.037	65	46258	4.810	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.200%
32) Benzene-d6	5.053	84	186723	4.157	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
36) 1,2-Dichloropropane-d6	6.072	67	58198	4.646	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.317	98	157634	3.781	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.600%
43) trans-1,3-Dichloroprop...	7.625	79	16046	3.811	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.200%
46) 2-Hexanone-d5	8.092	63	60080	40.654	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.300%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	33162	4.190	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.800%
66) 1,2-Dichlorobenzene-d4	11.622	152	52259	4.419	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.400%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.513	84	3264	0.244	ug/L	98
22) cis-1,2-Dichloroethene	3.921	96	3976	0.327	ug/L	98
30) Cyclohexane	4.683	56	7671	0.461	ug/L	92
33) Benzene	5.101	78	386845	8.435	ug/L	100
35) Methylcyclohexane	6.133	83	1550	0.080	ug/L #	70
52) Ethylbenzene	9.024	91	2275	0.043	ug/L	99
53) m,p-Xylene	9.149	106	2174	0.107	ug/L	74
60) Isopropylbenzene	9.934	105	2721	0.062	ug/L #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025565.D
Acq On : 15 Apr 2022 14:09
Operator : SY/MD
Sample : N2356-10DL 100X
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
GBDS8DL

Quant Time: Apr 16 02:42:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 02:38:36 2022
Response via : Initial Calibration

