

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025574.D
 Acq On : 15 Apr 2022 17:40
 Operator : SY/MD
 Sample : N2384-11
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6Q2

Quant Time: Apr 16 02:44:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 02:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	125764	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	123453	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	51489	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	43692	3.737	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	74.800%
7) Chloroethane-d5	1.574	69	43279	3.479	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	69.600%
11) 1,1-Dichloroethene-d2	2.114	63	67608	2.886	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	57.800%#
20) 2-Butanone-d5	3.902	46	69259	48.696	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.400%
24) Chloroform-d	4.352	84	75137	4.253	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
26) 1,2-Dichloroethane-d4	5.037	65	35241	4.629	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.600%
32) Benzene-d6	5.053	84	144508	4.046	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.000%
36) 1,2-Dichloropropane-d6	6.069	67	42989	4.315	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.400%
41) Toluene-d8	7.316	98	116900	3.527	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	70.600%
43) trans-1,3-Dichloroprop...	7.628	79	11491	3.432	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	68.600%
46) 2-Hexanone-d5	8.091	63	44255	37.659	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.320%
56) 1,1,2,2-Tetrachloroeth...	10.213	84	25830	4.104	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.000%
66) 1,2-Dichlorobenzene-d4	11.625	152	36146	4.187	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	83.800%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.133	85	107637	9.474	ug/L	98
13) Acetone	2.191	43	5056	5.237	ug/L	89
14) Carbon disulfide	2.301	76	9266	0.355	ug/L	97
22) cis-1,2-Dichloroethene	3.921	96	5010	0.521	ug/L	86
34) Trichloroethene	5.927	95	3122	0.322	ug/L	84
47) Tetrachloroethene	7.976	164	26882	3.546	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025574.D
Acq On : 15 Apr 2022 17:40
Operator : SY/MD
Sample : N2384-11
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6Q2

Quant Time: Apr 16 02:44:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 02:38:36 2022
Response via : Initial Calibration

