

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025579.D
 Acq On : 15 Apr 2022 19:37
 Operator : SY/MD
 Sample : N2384-19
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6R9

Quant Time: Apr 16 02:45:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 02:38:36 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	130404	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	133433	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	58357	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.311	65	46896	3.868	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	77.400%
7) Chloroethane-d5	1.571	69	47095	3.651	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.000%
11) 1,1-Dichloroethene-d2	2.114	63	73501	3.026	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.600%
20) 2-Butanone-d5	3.905	46	76619	51.954	ug/L	0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	103.900%
24) Chloroform-d	4.352	84	82578	4.507	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.200%
26) 1,2-Dichloroethane-d4	5.037	65	39795	5.041	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.800%
32) Benzene-d6	5.053	84	160598	4.160	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.200%
36) 1,2-Dichloropropane-d6	6.072	67	49829	4.628	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.600%
41) Toluene-d8	7.317	98	131025	3.657	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.200%
43) trans-1,3-Dichloroprop...	7.629	79	13363	3.692	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.800%
46) 2-Hexanone-d5	8.092	63	57436	45.220	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.440%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	29932	4.400	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.000%
66) 1,2-Dichlorobenzene-d4	11.622	152	44961	4.595	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.000%
Target Compounds						
					Qvalue	
34) Trichloroethene	5.921	95	8256	0.789	ug/L	94
35) Methylcyclohexane	6.130	83	1766	0.106	ug/L	92
42) Toluene	7.397	91	11439	0.267	ug/L	95
47) Tetrachloroethene	7.976	164	245268	29.935	ug/L	98
52) Ethylbenzene	9.018	91	7488	0.164	ug/L	99
53) m,p-Xylene	9.143	106	2137	0.123	ug/L	100
54) o-Xylene	9.551	106	970	0.057	ug/L	89
63) 1,2,4-Trimethylbenzene	10.915	105	2437	0.084	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025579.D
Acq On : 15 Apr 2022 19:37
Operator : SY/MD
Sample : N2384-19
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6R9

Quant Time: Apr 16 02:45:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 02:38:36 2022
Response via : Initial Calibration

