

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025584.D
 Acq On : 15 Apr 2022 22:44
 Operator : SY/MD
 Sample : N2438-01
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6R5

Quant Time: Apr 16 04:48:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 04:36:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.609	114	133667	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	119655	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.246	152	56022	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.298	65	59079	4.754	ug/L	-0.02
Spiked Amount	5.000	Range 40 - 130	Recovery	=	95.000%	
7) Chloroethane-d5	1.561	69	54446	4.118	ug/L	-0.01
Spiked Amount	5.000	Range 65 - 130	Recovery	=	82.400%	
11) 1,1-Dichloroethene-d2	2.102	63	86097	3.458	ug/L	-0.02
Spiked Amount	5.000	Range 60 - 125	Recovery	=	69.200%	
20) 2-Butanone-d5	3.976	46	133907	88.584	ug/L	0.08
Spiked Amount	50.000	Range 40 - 130	Recovery	=	177.160%#	
24) Chloroform-d	4.343	84	125002	6.657	ug/L	-0.01
Spiked Amount	5.000	Range 70 - 125	Recovery	=	133.200%#	
26) 1,2-Dichloroethane-d4	5.031	65	53878	6.659	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	133.200%#	
32) Benzene-d6	5.037	84	214382	6.192	ug/L	-0.02
Spiked Amount	5.000	Range 70 - 125	Recovery	=	123.800%	
36) 1,2-Dichloropropane-d6	6.069	67	45324	4.694	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	93.800%	
41) Toluene-d8	7.314	98	121120	3.770	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	75.400%	
43) trans-1,3-Dichloroprop...	7.625	79	11767	3.626	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	72.600%	
46) 2-Hexanone-d5	8.108	63	41334	36.290	ug/L	0.02
Spiked Amount	50.000	Range 45 - 130	Recovery	=	72.580%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	26528	4.349	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	87.000%	
66) 1,2-Dichlorobenzene-d4	11.622	152	44457	4.733	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	94.600%	
Target Compounds						
16) Methylene chloride	2.503	84	6852	0.610	ug/L	82
25) Chloroform	4.372	83	14368	0.827	ug/L	90
38) Bromodichloromethane	6.510	83	741	0.073	ug/L #	85
48) 2-Hexanone	8.162	43	10939	4.794	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025584.D
Acq On : 15 Apr 2022 22:44
Operator : SY/MD
Sample : N2438-01
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6R5

Quant Time: Apr 16 04:48:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 04:36:16 2022
Response via : Initial Calibration

