

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025587.D
 Acq On : 15 Apr 2022 23:54
 Operator : SY/MD
 Sample : N2438-05
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6S3

Quant Time: Apr 16 04:49:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 04:36:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	119119	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.850	117	123291	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	53998	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	41493	3.747	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	75.000%
7) Chloroethane-d5	1.571	69	45263	3.842	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.800%
11) 1,1-Dichloroethene-d2	2.111	63	67170	3.028	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	60.600%
20) 2-Butanone-d5	3.905	46	81896	60.794	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	121.580%
24) Chloroform-d	4.352	84	85741	5.123	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.400%
26) 1,2-Dichloroethane-d4	5.037	65	39772	5.516	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	110.400%
32) Benzene-d6	5.053	84	151676	4.252	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.000%
36) 1,2-Dichloropropane-d6	6.072	67	49662	4.992	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	99.800%
41) Toluene-d8	7.317	98	122525	3.701	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.000%
43) trans-1,3-Dichloroprop...	7.628	79	13062	3.906	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.200%
46) 2-Hexanone-d5	8.091	63	55321	47.138	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	94.280%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	33485	5.328	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	106.600%
66) 1,2-Dichlorobenzene-d4	11.622	152	42995	4.749	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025587.D
Acq On : 15 Apr 2022 23:54
Operator : SY/MD
Sample : N2438-05
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6S3

Quant Time: Apr 16 04:49:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 04:36:16 2022
Response via : Initial Calibration

