

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041522\
 Data File : VW025591.D
 Acq On : 16 Apr 2022 01:28
 Operator : SY/MD
 Sample : N2438-12
 Misc : 25mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW6S8

Quant Time: Apr 16 04:50:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 16 04:36:16 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	139773	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	143314	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.249	152	63309	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	51097	3.932	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	78.600%	
7) Chloroethane-d5	1.571	69	52613	3.806	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	76.200%	
11) 1,1-Dichloroethene-d2	2.111	63	78786	3.027	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	60.600%	
20) 2-Butanone-d5	3.899	46	94496	59.781	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	119.560%	
24) Chloroform-d	4.352	84	94353	4.805	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.000%	
26) 1,2-Dichloroethane-d4	5.037	65	46165	5.456	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	109.200%	
32) Benzene-d6	5.053	84	177769	4.287	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	85.800%	
36) 1,2-Dichloropropane-d6	6.069	67	56657	4.899	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	98.000%	
41) Toluene-d8	7.317	98	142301	3.698	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	74.000%	
43) trans-1,3-Dichloroprop...	7.625	79	15586	4.010	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	80.200%	
46) 2-Hexanone-d5	8.088	63	69665	51.067	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	102.140%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	38828	5.314	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	106.200%	
66) 1,2-Dichlorobenzene-d4	11.622	152	51388	4.841	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	96.800%	
Target Compounds						
					Qvalue	
13) Acetone	2.198	43	4565	4.255	ug/L	78
33) Benzene	5.114	78	3346	0.079	ug/L	100
42) Toluene	7.400	91	9679	0.210	ug/L	98
52) Ethylbenzene	9.017	91	4381	0.089	ug/L	98
53) m,p-Xylene	9.146	106	1461	0.078	ug/L	83
63) 1,2,4-Trimethylbenzene	10.918	105	1915	0.061	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041522\
Data File : VV025591.D
Acq On : 16 Apr 2022 01:28
Operator : SY/MD
Sample : N2438-12
Misc : 25mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW6S8

Quant Time: Apr 16 04:50:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040822WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 16 04:36:16 2022
Response via : Initial Calibration

